Biomechanical behavior of bone tissue in spine instrumentation

Comportamiento biomecánico de tejido óseo en fijaciones de columna

Armando Alpízar-Aguirre¹, Raide A. González-Carbonell^{2*}, Armando Ortiz-Prado³, Victor H. Jacobo-Armendáriz³, Osvaldo Ruiz-Cervantes³ y Lázaro Morales-Acosta³

¹Departamento de Cirugía de Columna, Instituto Nacional de Rehabilitación, Ciudad de México, México; ²Departamento de Ingeniería Mecánica, Facultad de Electromecánica, Universidad de Camagüey "Ignacio Agramonte Loynaz", Camagüey, Cuba; ³Departamento de Ingeniería de Diseño y Manufactura, Facultad de Ingeniería, Universidad Nacional Autónoma de México, Ciudad de México, México

Abstract

Background: The mechanical fixation of the spine in patients with osteoporotic vertebral degeneration is a challenge for surgeons, the vertebrae selected to insert the screws may fail, endangering health and even patient's life. **Objective:** The objective of the study was to study the effect of the variation of the bone density in the bone-screw interface from a three-dimensional model of the lumbar section. **Materials and methods:** The finite element method was used to model the behavior of the lumbar vertebral section when applying compression loads. **Results:** The stresses between 2 and 3 MPa were located on the contact surface with the screw, both in the vertebral body and in the apophysis, being slightly higher in the vertebral body. **Conclusions:** Regardless of bone density, the contact zones between the screws are susceptible to bone tissue failure. The posterior half of the vertebral body was the most sensitive to high values of stress, while in the areas furthest from the axis of the screw stress tended to their minimum.

Key Words: Osteoporosis. Pedicular fixation with screws. Stress. Biomechanics.

Resumen

Introducción: La fijación mecánica de la columna en pacientes con degeneración vertebral osteoporótica es un reto para los cirujanos, pues las vértebras seleccionadas para insertar los tornillos pueden fallar, poniendo en peligro la salud y la vida del paciente. **Objetivo:** Estudiar el efecto de la variación de la densidad ósea en la interfase hueso-tornillo, a partir de un modelo tridimensional de la sección lumbar. **Materiales y métodos:** Se emplea el método de los elementos finitos para modelar el comportamiento de la sección vertebral lumbar al aplicar cargas de compresión. **Resultados:** Los esfuerzos entre 2 y 3 MPa se ubicaron en la superficie de contacto con el tornillo, tanto en el cuerpo vertebral como en la apófisis, siendo ligeramente superiores en el cuerpo vertebral. **Conclusiones:** Independientemente de la densidad ósea, las zonas de contacto entre el tornillo son susceptibles al fallo del tejido óseo, debido a que están próximos al esfuerzo de fallo óseo de 2.37 ± 1.14 MPa reportado en la literatura. La mitad posterior del cuerpo vertebral fue la más sensible a sufrir valores altos de esfuerzos, mientras que en las zonas más alejadas del eje del tornillo los esfuerzos tendieron a su magnitud mínima.

Palabras Clave: Osteoporosis. Fijación pedicular con tornillos. Esfuerzos. Biomecánica.

Correspondencia:

*Raide A. González-Carbonell

Carretera Circunvalación, Norte km. 5½

C.P. 74650, Camagüey, Cuba

E-mail: raide.gonzalez@reduc.edu.cu

Fecha de recepción: 29-11-2018

Fecha de aceptación: 04-07-2019

DOI: 10.24875/CIRU.19000928

Cir Cir. 2020;88(1):41-48

Contents available at PubMed

www.cirugiaycirujanos.com

0009-7411/© 2019 Academia Mexicana de Cirugía. Publicado por Permanyer. Éste es un artículo open access bajo la licencia CC BY-NC-ND (<http://creativecommons.org/licenses/by-nc-nd/4.0/>).

Introduction

Osteoporosis is defined as the loss of trabecular bone which in the lumbar section of the spine can be present when the average bone mineral density is $< 0.8 \text{ g/cm}^2$. According to Martínez-Quiñones et al.¹, vertebral fractures of osteoporotic origin induce dorsal kyphosis, which can increase mortality by 23-35%, reduce respiratory vital capacity by 9% for each fractured dorsal vertebra, accelerate the sensation of fullness postprandial, and increase the degree of depression.

Patients with osteoporosis who have instability or deformities, fracture of vertebrae, or damage to the intervertebral discs need instrumentation in the spine as an orthopedic treatment. According to Burval et al.² with the increase in the patients' life expectancy, surgeries are having major challenges in the reconstruction of the spine due to the increase in bone fragility in advanced ages.

Mechanical fixation of the spine in patients with osteoporotic vertebral degeneration remains a challenge for surgeons³. In the middle of operations, the vertebrae selected to insert the screws may fail, endangering the patient's life. Methods of transpedicular fixation with screws offer stability in the spine and have proved to be the most rigid posterior fixation methods in the thoracic and lumbar part of the spine^{4,5}.

According to Alkaly and Bader⁶, despite the technological advances in the design of internal fixation systems, the loosening of the implants, the catastrophic failure in the bone-screw interface, the migration of material, and the loss of stability also constitute a serious complication in spine surgery in adults. It has been identified that the biomechanical factors that affect the fixation capacity of the pedicle screws are the design of the screws, the anatomical characteristics of the vertebral body and the pedicle (vertebrae geometry and the mechanical properties of the vertebral bone), and the techniques of insertion of the screw⁷.

The mechanical properties of the bone have a dependence on bone density, Seebeck et al.⁸ concluded that cortical tissue thickness and spongy bone density explain more than 90% of the variance of the final load of the screws under traction conditions and cantilever bending.

The conventional transpedicular screws and the cortical screws are two of the fixation techniques. Conventional transpedicular screws have a poor

performance due to the reduction of spongy bone mass that can severely limit the fixation potential in patients suffering from osteoporosis. On the other hand, the trajectory of insertion of the cortical screws theoretically maximizes the grip with the cortical tissue and potentially improves the long-term fixation.

Biomechanical analysis contributes to study the behavior of the human body under different loading conditions^{9,10}. In this direction, Sansur et al.³ were the first to perform a comparative study between conventional transpedicular fixation and cortical fixation following the physiological stresses in the osteoporotic lumbar spine. They found that the properties of transpedicular and cortical fixation were dependent on the level of the instrumented vertebra. Fixation with cortical screws exhibited a marked increase in the average load of failure in the lower lumbar segments. The transpedicular screw fixation showed an inverse tendency of lower failure loads in the lower vertebral elements and higher failure loads in the upper vertebral elements, probably secondary to the unique composition of the pedicle in the lumbar spine. On the other hand, Matsukawa et al.¹¹ investigated how the extraction force of the pedicle screw was affected by the insertion trajectory. They concluded that the effects of different trajectories were influenced by the geometry of the vertebra.

Despite the advances related to the spinal instrumentation, it has not been possible to confirm the study of the behavior of the bone-vertebra interface for different bone densities. It is necessary to study the effect of bone density variation in the bone-screw interface, from a three-dimensional (3D) model of the lumbar section.

Materials and methods

Geometry model of the study

This study focuses on the stresses that originate in the screw-bone interface for different bone densities. The Finite Element Method (FEM) is the tool for excellence used in modeling the behavior of bones subjected to the action of mechanical loads¹²⁻¹⁴. The analysis carried out with the FEM, starting with the simplification of the real object and the simplified object is called model¹⁵. In the present research, a patient-specific model was generated, where geometry consisted of three vertebrae in segment L1-L3 of the images of the free software library 3D Slicer 4.8.0 (<http://www.slicer.org>)¹⁶. The medical images information was provided

in the nearly raw raster data (NRRD) format which is a file format designed to support scientific visualization and to process images involving N-dimensional raster data¹⁷.

Once the images were imported into the 3D Slicer, we proceeded to render the data of the file in NRRD, according to the Volume Rendering module. The area of interest (region of interest) was selected and the work area was delimited with the Crop Volume tool. Segmentation of the images was performed with the Segment Editor module, where a layer was created for the bone tissue, which allowed to manually edit the groups of cuts and thus to eliminate surface imperfections, such as holes inside the solid, and avoid future errors in the mesh and in the analysis by MEF.

With the tool set surface smoothing, the smoothing of the geometric irregularities due to the segmented imperfections was performed. Smoothing prevents the generation of the surface mesh with triangles of poor quality. Once the process of editing the images was finished, the changes made to the model in STL extension were saved.

Subsequently, the geometry was imported into the Autodesk Meshmixer 3.3.15 software (Autodesk, Inc., CA, USA) to perform a new smoothing of the surface to eliminate the imperfections that were not possible to be eliminated in the 3DSlicer. In this same software, we converted the surface mesh into a solid element.

To separate the vertebrae and to insert the pedicle screws, the geometry was imported into the SolidWorks Premium 2018 software (Dassault Systemes SOLIDWORKS Corp., MA, USA). In this software, the starting and ending points of the screw trajectory were created, represented by the blue axes of figure 1A and 1B. The trajectory is parallel to the plates of the vertebral body (Fig. 1A). The holes of the screws were considered as a smooth of constant diameter. Reference planes were created perpendicular to the path of the screws for that portion of the vertebra volume that was subtracted. The geometry of the screw was created in a separate file, a diameter of 5.5 mm and a length of 30 mm were defined in the area of interaction with the vertebra.

Then, the assembly operation of the screws and the vertebra was performed, using the concentricity restriction between the diameter of the screws and the holes of the vertebrae (Fig. 1C). This geometry was exported in a file with Parasolid x_t extension (Siemens PLM Software, TX, USA).

With the geometry of the vertebrae processed, we proceeded to define the geometry model to be used

Figure 1. Geometric model of L3 vertebrae with cortical pedicle screws. **A:** insertional trajectory of screws in sagittal plane view. **B:** in transversal plane view. **C:** assembly of screws inserted into the vertebrae.

in the simulation by the MEF. This simulation was carried out in the Abaqus CAE 6.14 software (Dassault Systemes SIMULIA Corp., MA, USA). Geometry was imported into the Abaqus. In the Part module, the work surfaces that will be used during the definition of the model were created. The geometry model carried the information about the positioning of the elements of the assembly, but it was necessary to define the interaction between the work surfaces.

Model of mechanical properties of the material

The mechanical properties of the bone can be considered as “unique” since they not only depend on the tissue being analyzed but also on the characteristics

Figure 2. Model of loads and constraints of the mechanical fixation of a vertebral segment. **A:** the hardware releases the intervertebral disc of its compression loading and the force is transmitted from screw to screw, from the upper vertebra to the lower ones through the hardware. All degrees of freedom of the screw head were removed and a compression load of 500 N was applied to the vertebra plate such as. **B:** upper plate for upper vertebra. **C:** bottom plate for the lower vertebra.

of the individual, the conditions to develop the mechanical tests, the preparation of the sample, and the speed of deformation. In addition, the bone is able to adapt to the loading conditions, modifying its structure because there is a large dispersion in the mechanical properties reported in literature^{18,19}.

In literature, there is a great diversity in terms of the mechanical properties of a spongy tissue. The configuration of the trabeculae changes in each type of bone and within the same bone²⁰⁻²². It is suggested that the modulus of elasticity (E) of the vertebral tissue is approximately 1 GPa, the great dispersion of values in this tissue is influenced not only by the stiffness of the trabecula but also by its orientation or disposition²³⁻²⁵. Banse et al.²⁶ limited the yield stress range between 0.6 and 6.17 MPa.

In the definition of the mechanical properties of bone tissue, the relation between the apparent density and E (1) was used, as that expression E can range

from 920 MPa ($p = 0.35$) to 100 MPa ($p = 0.09$), the Poisson's coefficient was set as 0.2. For each simulation, different homogenous values of E were assigned throughout the vertebra and it was considered as a material with isotropic behavior²⁶⁻²⁸.

$$E = 4750 \cdot \rho^{1.56} \quad (1)$$

Model of loads and constraints

In the mechanical fixation of a vertebral segment, the hardware releases the intervertebral disc of its compression loading, so it can be assumed that the upper vertebra of the instrumented segment receives the load of the adjacent segments by the upper plate, which is transmitted from screw to screw, from the upper vertebra to the lower ones through the hardware, and finally, the load is transmitted to the adjacent segment by the bottom plate of the lower vertebra (Fig. 2A).

Figure 3. A-B: behavior of maximum von Mises stresses for bone tissue.

To simulate the behavior of the upper vertebra, a vertical compression load of 500 N was applied on the upper plate, which takes into account the effect of the paraspinal muscles and the intra-abdominal pressure in the balanced foot position^{27,28}. A set of nodes was defined on the surface of the upper plate, which was selected as a reference to apply the concentrated load. The movement on the X, Y, and Z axes was eliminated on the outside face of the screw head (Fig. 2B).

To simulate the behavior of the lower vertebra, a vertical compression load of 500 N was applied to the bottom plate. A set of nodes was also defined on the surface of the bottom plate, which was selected as a reference to apply the concentrated load. The movement in the X, Y, and Z axes was eliminated on the external face of the head of the screws (Fig. 2C).

Geometry mesh

Subsequently, the geometric shape of the bone in the analysis is approximated by the division of the continuous solid into finite elements. This process is called discretization or meshing of the continuum. The elements are joined through the nodes and make up the mesh. The FEM is a numerical method and, therefore, its solution is approximate. The magnitude of the error depends on the correct definition of the model and the quality of the mesh.

Both the screw and the vertebra were meshed using tetrahedral elements with an average size of 1.6 mm, following a free mesh and using the advancing-front technique (AFT) and the Delaunay algorithm. The mesh of the vertebra contains 247,631 elements, while that of each screw 497 elements.

Results and discussion

During the analysis of a 3D object subjected to loads, the reactions that appear inside it are opposed

to external forces. This way, stress within the body arises (force density per unit area). If an infinitesimal element is considered inside the cube-shaped body, normal and tangential stresses appear on each side. Through one of the resistance theories, the equivalent stress of each infinitesimal element is obtained. To determine these stresses, the theory of the potential energy of deformation or also known as von Mises was selected. The interpretation of the results was based on the maximum equivalent von Mises stresses, expressed in MPa.

In the analysis, two mechanical properties were defined for each material, one for the screw and one for the bone, and for the load condition to which both parts of the assembly are subject, the maximum stresses were much higher in the screw than in the vertebra, but the metallic body has greater strength than the bone so that bone failure is more likely to occur. In this sense, the maximum scale of the color gradient of the graph was redefined to highlight the stresses below 10 MPa since the stresses in the vertebra are less than this magnitude and it is also higher to the yield stress reported by the trabecular bone of the vertebrae (Fig. 3)^{21,29,30}.

When the instrumentation of a vertebral section is performed, one vertebra receives the load on its upper plane and the other on the lower plane. The load is transmitted from one vertebra to the other through the screws and rods, leaving the intervertebral disc free from bearing loads. Fig. 3 shows the intensity of loads of the bone tissue of the vertebra which supports load on the upper plate. The highest values (3-5 MPa) were reported in the area where the screw hole ends, which is a sudden change of section, so this result can be considered with caution. In real conditions, this change of section does not exist since the hole ends in a point.

The stress in the order of 2-3 MPa was located on the surface of contact between bone and screw, located both in the vertebral body and in the apophysis,

Figure 4. Behavior of maximum von Mises stresses in the inner of the vertebra when it was loaded on the upper plate.

Figure 5. Behavior of maximum von Mises stresses in the inner of the vertebra when it was loaded on the bottom plate.

being slightly higher in the vertebral body and there was a tendency to decrease as it moves away from the vertebral body. The posterior half of the vertebral body was the most sensitive to high values of stress, while in the furthest areas from the axis of the screw, the stresses tended to their minimum magnitude (Fig. 4). There is a correspondence with the result obtained by Xu et al.²⁸

Fig. 5 shows the behavior of the stresses in the bone tissue of the vertebra supporting the lower plate. As shown, there is a behavior similar to that obtained when the vertebra is subjected to compression loads in the upper plate. Again, the stresses in the order of 2-3 MPa are in the area of contact of the screw with

the bone. There is a slight difference in the distribution of stresses. In the vertebrae with compression load in the upper plate, the amount of elements above the screw with the stresses is bigger than when the load is applied in the lower part and vice versa.

A load of 500 N was applied to the models analyzed. From the result obtained, it is recommended for the patient not to carry out heavy activities since a greater demand on the column results in an increase in stress, thus with an increase in the stress in the bone. According to the results of the stress analysis, the areas of contact between the screw are susceptible to bone tissue failure because they are close to the bone failure stress of 2.37 ± 1.14 MPa reported by

Banse et al.²⁶ These authors carried out the characterization of the mechanical properties and they discovered that the vertebral body can resist up to 6.17 MPa, but in the areas, where the bone density is close to 0.35 g/cm³.

The common problems of the implants are the loosening of the screw-bone interface, extraction, and insufficiency of the instruments after the lumbar arthrodesis (the catastrophic screw or hardware failure)⁶. Product of the working circumstances of the clamp, when the spine is bent forward, the screws tend to come out of the bone. On the other hand, due to the direction of the loads caused by the patient's weight, the screws can fail at shear loads. One aspect to take into account is the cyclical nature of the loads caused by the gear, which brings about the risk of failure due to fatigue, mainly in the concentration zones of stress located in the threads of the threads³¹.

As limitations of the model, only the compression loads in the vertebral body were considered in the present study. According to Matsukawa et al.¹¹, the load in axial direction to the axis of the screw that can resist the bone before the loosening occurs is $1,040 \pm 268$ N. Although with a marked effect on the screws, the fatigue caused by the cyclic nature of the loads brought about the march may also be present in the bone tissue and it can affect the screw-bone interaction, causing the non-resistance of the bone to the loads reported by Matsukawa et al.¹¹

For future analysis, adjacent segments and intervertebral discs should be included, although only the screw-bone interface was studied, other researchers suggested that spinal fixation is a possible cause of degeneration of adjacent disc segments. It has been suggested that mechanical loads and stresses alter the structure and the failure properties of the disc, which eventually leads to the degeneration of the intervertebral discs. In addition, the greater rigidity induced by the pedicle screws could be attributed to the main function they perform to restrict the fused joint during flexion-extension loading conditions. To counteract the damage in discs of the adjacent segments, it is decided to reduce the rigidity of the instrumented segment.

According to Elmasry's study³², the mobility and stresses experienced by a spinal segment increase when neighboring segments are fused. Therefore, control of the stresses generated in the adjacent segment due to the implanted construction provides an indication of the potential risks of degeneration in the adjacent intervertebral disc.

According to the study of mobility of the intervertebral discs in posterior pedicle fixation performed by Rohlmann et al.²⁷, to achieve a natural intervertebral movement, an elastic modulus of the low rod is required, while to reduce the intradiscal pressure, a high rod rigidity is required. On the other hand, to reduce the forces of the facet joint, it is necessary that the connection between the rod and the pedicle screw is rigid. Hence, it can be deduced that the degree of mobility of the instrumented spinal segment will be sacrificed.

Conclusions

According to the analysis of stress distribution, the zones of contact between the screw and the bone are susceptible to bone tissue failure because they are close to the stress of bone failure of 2.37 ± 1.14 MPa reported in literature. The posterior half of the vertebral body was the most sensitive to high values of stress, while in the areas furthest from the axis of the screw stresses tended to their minimum magnitude. From the obtained result, it is recommended that the patient cannot carry out heavy activities because a higher demand on the spine results in an increase in strength, therefore, an increase in the stress inside the bone.

Acknowledgment

This research work was carried out to thank the support of a scholarship of excellence granted by the Government of Mexico, through the Secretaría de Relaciones Exteriores.

This research was also supported by the Project "PAPIIT IN115415 Validación experimental de modelos de tejido óseo desarrollados en elemento finito" which was supported by Dirección General de Asuntos del Personal Académico of the Universidad Nacional Autónoma de México. The authors acknowledge the financial support of these institutions and we thank to the Instituto Nacional de Rehabilitación of the Secretaría de Salud Pública de México for providing medical images and the advice of its experts on Orthopedics topics.

Conflicts of interest

The authors declare that they have no conflicts of interest relevant to this article.

Ethical disclosures

Protection of human and animal subjects. The authors declare that no experiments were performed on humans or animals for this study.

Confidentiality of data. The authors declare that no patient data appear in this article.

Right to privacy and informed consent. The authors declare that no patient data appear in this article.

References

- Martínez-Quiriones JV, Aso-Escario J, Arregui-Calvo R. Refuerzo vertebral percutáneo: vertebroplastia y cifoplastia. *Procedimiento técnico*. Neurocir. 2005;16:427-40.
- Burval DJ, McLain RF, Milks R, Inceoglu S. Primary pedicle screw augmentation in osteoporotic lumbar vertebrae: biomechanical analysis of pedicle fixation strength. *Spine (Phila Pa 1976)*. 2007;32:1077-83.
- Sansur CA, Caffes NM, Ibrahim DM, Pratt NL, Lewis EM, Murgatroyd AA, et al. Biomechanical fixation properties of cortical versus transpedicular screws in the osteoporotic lumbar spine: an *in vitro* human cadaveric model. *J Neurosurg Spine*. 2016;25:467-76.
- Abumi K, Itoh H, Taneichi H, Kaneda K. Transpedicular screw fixation for traumatic lesions of the middle and lower cervical spine: description of the techniques and preliminary report. *J Spinal Disord*. 1994;7:19-28.
- Moldavsky M, Salloum K, Bucklen B, Khalil S, Mehta JS. *In vitro* biomechanical study of pedicle screw pull-out strength based on different screw path preparation techniques. *Indian J Orthop*. 2016;50:177-82.
- Alkaly RN, Bader DL. The effect of transpedicular screw design on its performance in vertebral bone under tensile loads: a parametric study. *Clin Spine Surg*. 2016;29:433-40.
- Tsuang FY, Chen CH, Wu LC, Kuo YJ, Lin SC, Chiang CJ, et al. Biomechanical arrangement of threaded and unthreaded portions providing holding power of transpedicular screw fixation. *Clin Biomech (Bristol, Avon)*. 2016;39:71-6.
- Seebeck J, Goldhahn J, Städele H, Messmer P, Morlock MM, Schneider E, et al. Effect of cortical thickness and cancellous bone density on the holding strength of internal fixator screws. *J Orthop Res*. 2004;22:1237-42.
- González-Carbonell RA, Ortiz-Prado A, Jacobo-Armendáriz VH, Cisneros-Hidalgo YA, Alpízar-Aguirre A 3D patient-specific model of the tibia from CT for orthopedic use. *J Orthop*. 2015;12:11-6.
- Carbonell RA, Padrón EN, Pérez BC, Hidalgo YC, Sorí ML. Carácter interdisciplinario de la modelación computacional en la solución de problemas de salud. *Hum Méd*. 2014;14:646-58.
- Matsukawa K, Yato Y, Hynes RA, Imabayashi H, Hosogane N, Yoshihara Y, et al. Comparison of pedicle screw fixation strength among different transpedicular trajectories: a finite element study. *Clin Spine Surg*. 2017;30:301-7.
- Hidalgo YA, Carbonell RA, Alvarez AP, Corona EC, Rodríguez YO. Generación de imágenes tridimensionales: integración de tomografía computarizada y método de los elementos finitos. *Rev Cubana Inv Bioméd*. 2014;33:313-21.
- Carbonell RA, Prado AO, Armendáriz VH, Hidalgo YA, Acosta LM. Consideraciones en la definición del modelo específico al paciente de la tibia. *Rev Cubana Inv Bioméd*. 2015;34:157-67.
- González-Carbonell RA, Ortiz-Prado A, Cisneros-Hidalgo YA, Alpízar-Aguirre A. Bone remodeling simulation of subject-specific model of tibia under torque. In: Braidot A, Hadad A, editors. VI Latin American Congress on Biomedical Engineering CLAIB 2014, Paraná, Argentina 29, 30 and 31 October 2014. IFMBE Proceedings, No. 49. Berlin: Springer International Publishing; 2015. p. 305-8.
- González Carbonell RA, García EA, Pérez YC. Tacón de torque. Análisis tensional y deformacional utilizando el método de elementos Finitos. *Ing Mecánica*. 2007;10:79-83.
- Fedorov A, Beichel R, Kalpathy-Cramer J, Finet J, Fillion-Robin JC, Pujol S, et al 3D slicer as an image computing platform for the quantitative imaging network. *Magn Reson Imaging*. 2012;30:1323-41.
- Hosny A, Shen T, Kuo AS, Long D, Andrawes MN, Dilley JD, et al. Unlocking vendor-specific tags: three-dimensional printing of echocardiographic data sets. *J Thorac Cardiovasc Surg*. 2018;155:143-50.
- Keaveny TM, Morgan EF, Yeh OC. Bone mechanics. In: Standard Handbook of Biomedical Engineering and Design. Ch. 8. New York: McGraw-Hill Handbooks; 2004. Available from: http://www.server0.unhas.ac.id/tahir/BAHAN-KULIAH/BIO-MEDICAL/NEW/HANBOOK/Bone_Mechanics.pdf.
- Meyer U, Handschel J. Influence of biomechanical loads. In: Meyer U, Handschel J, Wiesmann HP, Meyer T, editors. Fundamentals of Tissue Engineering and Regenerative Medicine. Berlin, Heidelberg: Springer; 2009. p. 705-17.
- Rho JY, Kuhn-Spearing L, Zioupos P. Mechanical properties and the hierarchical structure of bone. *Med Eng Phys*. 1998;20:92-102.
- Morgan EF, Keaveny TM. Dependence of yield strain of human trabecular bone on anatomic site. *J Biomech*. 2001;34:569-77.
- Keaveny TM, Morgan EF, Niebur GL, Yeh OC. Biomechanics of trabecular bone. *Annu Rev Biomed Eng*. 2001;3:307-33.
- Ruiz O, Schouwenaars R, Ramírez EI, Jacobo VH, Ortiz A, editors. Analysis of the Architecture and Mechanical Properties of Cancellous Bone Using 2D Voronoi Cell Based Models. Proceedings of the World Congress on Engineering; 2010.
- Ruiz O, Schouwenaars R, Ramírez EI, Jacobo VH, Ortiz A. Effects of architecture, density and connectivity on the properties of trabecular bone: a two-dimensional, Voronoi cell based model study. *AIP Conf Proc*. 2011;1394:77-89.
- Carretta R, Lorenzetti S, Müller R. Towards patient-specific material modeling of trabecular bonepost-yield behavior. *Int J Numer Methods Biomed Eng*. 2013;29:250-72.
- Banase X, Sims TJ, Bailey AJ. Mechanical properties of adult vertebral cancellous bone: correlation with collagen intermolecular cross-links. *J Bone Miner Res*. 2002;17:1621-8.
- Rohlmann A, Nabil Boustani H, Bergmann G, Zander T. Effect of a pedicle-screw-based motion preservation system on lumbar spine biomechanics: a probabilistic finite element study with subsequent sensitivity analysis. *J Biomech*. 2010;43:2963-9.
- Xu G, Fu X, Du C, Ma J, Li Z, Ma X, et al. Biomechanical effects of vertebroplasty on thoracolumbar burst fracture with transpedicular fixation: a finite element model analysis. *Orthop Traumatol Surg Res*. 2014;100:379-83.
- Chen Y, Ma HT, Liang L, Zhang C, Griffith JF, Leung PC, editors. A Simulation Study on Marrow Fat Effect on Biomechanics of Vertebra Bone. 2015 37th Annual International Conference of the IEEE Engineering in Medicine and Biology Society; 2015.
- Wang J, Zhou B, Liu XS, Fields AJ, Sanyal A, Shi X, et al. Trabecular plates and rods determine elastic modulus and yield strength of human trabecular bone. *Bone*. 2015;72:71-80.
- Giacaglia GE, Lamas WD. Pedicle screw rupture: a case study. *Case Stud Eng Fail Anal*. 2015;4:64-75.
- Elmasry S, Asfour S, Travascio F. Implications of spine fixation on the adjacent lumbar levels for surgical treatment of thoracolumbar burst fractures: a finite element analysis. *J Spine Care*. 2016;1:1-5.